

Arise® 1000 EX

Low Air-Loss Bariatric Surface

Rental Solutions Delivered by



1 800 STRYKER Option 3

Universal | Hospital | Services®



Design and aesthetics allow for flexibility to serve as a standard hospital bed surface

Inflatable width and length air bladders can be customized to each individual patient

Twenty individual therapy cells help to evenly distribute the patient's weight and maximize pressure relief and ensure patient comfort

Pulsating feature oscillates air throughout the mattress every 30 seconds providing increased patient comfort

Foam bottom ensures that patients will never bottom out

CairRails™ provide an additional two inches of patient support

Flexible Design for Multi-Purpose Use

The Stryker Arise 1000 EX surface design and aesthetics provide versatile options for all patients. The surface can be customized to each patient with intuitive controls, and inflatable width and length air bladders. The combination of low air-loss and pulsation therapy provide the best in patient care and wound healing. To maximize patient comfort during transports, the visco-elastic foam bottom ensures that patients will never bottom out.



Customize surface width and length dimensions with intuitive controls

Arise 1000 EX

Low Air-Loss Bariatric Surface

stryker®



Medical

3800 E. Centre Ave.
Portage, MI 49002 USA
t: 269 329 2100 f: 269 329 2311
toll free: 800 787 9537

www.stryker.com

Standard Features

- Inflatable width air bladders
- Inflatable length air bladders
- Low air-loss
- Maximum inflate
- Pulsation therapy
- Surface comfort adjustment
- CairRails
- Foam crib bottom
- Nylon cover
- Therapy control unit

Mattress Specifications

• Model Number	2236
• Overall Length	80, 88" (233.2, 223.5 cm)
• Overall Width	36, 42, 48" (91.4, 106.7, 121.9 cm)
• Thickness	8.5" (22 cm)
• Weight Capacity	1000 lbs (453.6 kg)

Electronics

• Agency Approvals	Medical Equipment, UL 60601-1; CAN/CSA C22-2 No. 601-1
• Ampere Rating	5A
• Volts	120V AC
• Current Leakage	< 100 Microamperes
• Frequency	60 Hz
• Hospital Grade Plug	Yes

*Stryker reserves the right to change specifications without notice.
Sleep surfaces are manufactured for Stryker Medical.*

Copyright © 2008 Stryker
PC/SCG-7.5M-0508, 07-SMD-2184
Mkt Lit-216-122107 Rev B



Mixed Sources
Product group from well-managed
forests, controlled sources and
recycled wood or fibre
Cert no. SW-COC-1530
www.fsc.org
© 1996 Forest Stewardship Council